



Mini-Review on Mini-Metal Implants for Focal Cartilage Lesions

Johannes Holz, Ansgar Ilg, Rene Kaiser and Stefan Schneider

Ortho Centrum Hamburg, Germany

Mini Review

Recently, two studies have been published on a patient specific metal implant for focal chondral lesions in the knee where clinical results as well as revision data were quite favorable [1,2]. In the quest for an acceptable mode of treatment for the so-called GAP-patient [3], i.e. patients that are too old for biologic treatment and too young for joint replacements, these reports might indicate a new level of success and possibly an increasing interest for such mini-metal implants.

Focal lesions of cartilage in the knee is generally viewed as a common precursor of osteoarthritis (OA) [4,5] and in view of the current epidemic of knee OA [6], any attempt at reducing the need for TKA would appear warranted. Moreover, with increasing demands on functionality, up to 20% of patients receiving a TKA express discontent with the procedure [7,8]. In the last 25 years, intensive research has been carried out on biologic methods of transplantation of chondrocytes into these focal defects [9-11]. These attempts have been successful to an extent but are generally limited to younger patients, under 30 to 35 years of age. Also, biologic methods are plagued with long rehab periods of 12 to 18 months [12] and have been subject to regulatory constraints. Therefore, the various alternatives to chondrocyte transplantation have not met the success that was originally anticipated.

A hard resurfacing implant, where the rehab period is limited to soft-tissue healing time, may appear beneficial. Such hard implants have appeared during the last decade [13] and have shown promising results [14-17]. They all consist of round or elongated hats with one or two fixation peg(s) of some particular design. They are all aiming at the distal femur and are of the uni-polar design, i.e. the implant material articulates directly against normal hyaline cartilage on the tibia or patella (although the HemiCap can be used with a patellar implant). This is a short review of relevant issues regarding these implants to preserve the knee joint from knee replacements in the middle-aged population.

A first-generation implant (the HemiCAP family, Arthrosurface, US) appeared in the mid 2000s and consists of a CoCr metal alloy hat connected to a titanium fixation screw, joined by a Morse-taper. The articulating hat comes in an off-the-shelf library with a number of different shapes where the best is fitted to the particular joint using so called sizing cards. The system uses a guide to fit a Kirschner wire perpendicular to all tangents of the cartilage surface to be replaced as the base for implantation. A number of reports have shown good clinical results [14,18,19]. Indeed, a case-report on 2 cases after 12 years showed good clinical results and no deterioration of the opposing cartilage [20]. However, a number of studies show disconcerting revision figures in the order of 25% after 5 years [21,22].

Somewhat later, another hard-material implant, the BioPoly (BioPoly, US) [16], appeared. This is a one-piece titanium implant with a polyethylene cap towards the joint cavity. The polyethylene is enhanced ("micro-composite") with hyaluronic acid in order to be gentler to the opposing cartilage. There is one published report on 40 patients of which 12 were followed over the full 2-years FU-period [16]. Initial results were reported to be good with clinical improvement of the same order of magnitude as other treatment modalities and only one case had been revised within this short period of time.

At about the same time, a third metal implant for focal chondral lesions appeared, the Episealer implant (Episurf Medical, Sweden). This technology is based on three fundamentals; a MRI-based Damage Marking Report (DMR), individualized guide instruments based on the MRIs and an individualized implant again from the MRIs [23]. The implant itself is a one-piece CoCr alloy hat with one or two fixation peg(s). Surfaces meeting bone/cartilage are coated with a double coating of

OPEN ACCESS

*Correspondence:

Johannes Holz, Ortho Centrum
Hamburg, Hansastrasse 1-3, 20149
Hamburg, Germany,
E-mail: dr.holz@oc-h.de

Received Date: 24 Dec 2020

Accepted Date: 22 Jan 2021

Published Date: 09 Feb 2021

Citation:

Holz J, Ilg A, Kaiser R, Schneider S.
Mini-Review on Mini-Metal Implants
for Focal Cartilage Lesions. *Clin Surg*.
2021; 6: 3056.

Copyright © 2021 Johannes Holz. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

titanium and hydroxyapatite for osteo- and chondrointegration [24]. Recent published papers report good clinical results [1] and favorable revision figures in mid-term FU [2].

Opposing Cartilage

Arguably the most important issue with these uni-polar implants is how the opposing cartilage reacts to the artificial material. Here, this natural tissue will meet artificial metal/polymers under the extremely harsh mechanical conditions that prevail in the human knee where compressive forces of 4x to 6x body weight occur. There are some previous experiences. Normal joint cartilage articulates with patellar cartilage in TKA where the patella is not resurfaced and this appears to go well [25]. However, uni-polar hip replacement has a mixed track record [26,27]. These focal implants also show mixed records. The first-generation implant has shown conspicuous signs of wear of the opposing cartilage in some studies [21,28] while the very same implant explicitly has shown little problem of this kind with an unchanged medial joint space after 5 years [14]. Again, after as long as 12 years, the opposing cartilage was reported to hold up nicely. This may suggest that it is not the material itself but rather the precision of insertion and preoperative indications that are crucial. Obviously, a metal implant must not sit proud or it will harvest the opposing cartilage [29,30]. Therefore, precise guide instruments may be of decisive importance. The Episealer MRI-based individualized guides allow for incremental fine-tuning of insertion depth in steps of 200 µm to consistently position the implant about 0.5 mm below the surrounding cartilage and this position is checked with a “dummy”, an exact replica of the implant. Human cartilage asymptotically compresses by about 20% [31] and 0.5 mm, corresponds to 20% of the cartilage thickness (one-sided) in a normal knee [32].

Surrounding Cartilage

The surrounding cartilage must accept and preferably integrate with the artificial material. From hip- and knee replacements, it is well-known how wear products from joint replacements can invade the peri-prosthetic space and create osteolysis [33]. It is improbable that the materials in these focal implants should wear into malignant particles but even joint fluid itself is aggressive and could cause osteolysis [34,35]. For this reason, bonding of cartilage to implant suggests a feasible construct. Of the three focal implants, only the Episealer features a provision for bonding to cartilage. The HA-coating of the periphery of the hat facing cartilage has been shown to produce so called chondrointegration, i.e. a bond between implant and cartilage effectively sealing the cartilage defect [24].

Moreover, and contrary to any other mode of treatment, these mini-metal implants immediately support the surrounding cartilage. Edge loading of the cartilage rim surrounding the lesion is subject to excessive stress on normal locomotion. This creates a progressive increase in size of the lesion, the so-called “pot hole” effect [36,37] whereby the focal lesion deteriorates into a progressively larger one, eventually leading to full-blown OA. With physical support of the cartilage rim, excessive stress/strain is counteracted and may well be a decisive factor for better longevity.

Bonding

All three focal implants are inserted without bone cement but rely on direct bonding to bone. The HemiCAP and the BioPoly implants bond to bone by osseointegration to titanium [38,39], a time-honored, consistent and well documented mode of fixation. Despite

this, the Australian registry reports a substantial number of loosening for the HemiCAP, presumably being caused by failure of the Morse taper. Technically, the insertion of the titanium screw is an exacting procedure, not too deep or the Morse taper will not catch and not too shallow or the hat will protrude. The Episealer bonds by way of a double coating, hydroxyapatite superficially for fast bonding [40,41] and titanium underneath for consolidation of the fixation should there be hydrolysis of the hydroxyapatite [40]. No loosening has been reported for the Episealer device.

Patient Selection

As with any kind of surgical procedure, patient selection is crucial. For these patients, the problem often starts with a traumatic incident [42,43] that hits the cartilage often close to the apex of the femoral condyle. Sometimes this leads to an osteochondral fracture and a loose piece of cartilage in the knee. More often, however, the mechanical trauma is smaller having only a crushing effect of the intact cartilage at the apex. This may alter the metabolism of the chondrocytes into a more destructive pathway and the scene is set for a slow, continuous process towards OA [44]. Somewhere along this continuum, there is a window of opportunity for focal implants, when the lesion is not too large, when the opposing cartilage is not yet too damaged and when there is still a possibility to regain joint homeostasis [45]. This window can be sought by MRI that can delineate the extent of the lesion and the status of the subchondral bone as well as the status of opposing cartilage. For this reason, MRI is an integral part of the Episealer system and the DMR allows a careful selection of patient that is fit for the procedure. Possibly, too wide indications explain the untoward revision statistics for the HemiCAP [21,28].

MRI also allows examination of possible Bone Marrow Lesions (BML). Cartilage does not carry nerves, but subchondral bone does. Pain sits in the bone and large BMLs, or bony defects too, can be addressed by a thicker Episealer implant.

Conclusion

Small, metallic implants have the advantage of immediate function once soft-tissue healing after surgery has taken place. Post-operative rehabilitation is short. Recently published results indicate good clinical effects in terms of well-being and, provided correct patient selection and precise surgery, longevity appears to be consistent and of promising magnitude.

References

- Holz J, Spalding T, Boutefnouchet T, Emans P, Eriksson K, Brittberg M. Patient specific metal implants for focal chondral lesions in the knee; excellent clinical results at 2 years. *Knee Surg Sports Trauma Arthroscopy*. 2020.
- Martinez-Carranza N, Rockborn P, Roberts D, Högrström M, Stålmán A. Successful treatment of femoral chondral lesions with a novel customized metal implant at midterm follow-up. *Cartilage*. 2020;1947603520967064.
- Li CS, Karlsson J, Winemaker M, Sancheti P, Bhandari M. Orthopedic surgeons feel that there is a treatment gap in management of early OA: International survey. *Knee Surg Sports Traumatol Arthrosc*. 2014;22:363-78.
- Cicuttini F, Ding C, Wluka A, Davis S, Ebeling PR, Jones G. Association of cartilage defects with loss of knee cartilage in healthy, middle-age adults: A prospective study. *Arthritis Rheum*. 2005;52(7):2033-9.
- Schinhan M, Gruber M, Vavken P, Dorotka R, Samouh L, Chiari C. Critical-size defect induces unicompartmental osteoarthritis in a stable

- ovine knee. *J Orthop Res.* 2012;30(2):214-20.
6. Kurtz S, Ong K, Lau E, Mowat F, Halpern M. Projections of primary and revision hip and knee arthroplasty in the United States from 2005 to 2030. *J Bone Joint Surg Am.* 2007;89(4):780-5.
 7. Bourne RB, Chesworth BM, Davis AM, Mahomed NN, Charron KDJ. Patient satisfaction after total knee arthroplasty: Who is satisfied and who is not? *Clin Orthop Relat Res.* 2010;468(1):57-63.
 8. Parvizi J, Nunley RM, Berend KR, Lombardi AV Jr, Ruh EL, Clohisy JC. High level of residual symptoms in young patients after total knee arthroplasty. *Clin Orthop Relat Res.* 2014;472(1):133-7.
 9. Biant LC, Bentley G, Vijayan S, Skinner JA, Carrington RWJ. Long-term results of autologous chondrocyte implantation in the knee for chronic chondral and osteochondral defects. *Am J Sports Med.* 2014;42(9):2178-83.
 10. Brittberg M. Cell carriers as the next generation of cell therapy for cartilage repair: A review of the matrix-induced autologous chondrocyte implantation procedure. *Am J Sports Med.* 2010;38(6):1259-71.
 11. Minas T, Keudell AV, Bryant T, Gomoll AH. The John Insall Award: A minimum 10-year outcome study of autologous chondrocyte implantation. *Clin Orthop Relat Res.* 2014;472(1):41-51.
 12. Nho SJ, Pensak MJ, Seigerman DA, Cole BJ. Rehabilitation after autologous chondrocyte implantation in athletes. *Clin Sports Med.* 2010;29(2):267-82.
 13. Brennan SA, Devitt BM, O'Neill CJ, Nicholson P. Focal femoral condyle resurfacing. *Bone Joint J.* 2013;95-B(3):301-4.
 14. Becher C, Kalbe C, Thermann H, Paessler HH, Laprell H, Kaiser T. Minimum 5-year results of focal articular prosthetic resurfacing for the treatment of full-thickness articular cartilage defects in the knee. *Arch Orthop Trauma Surg.* 2011;131(8):1135-43.
 15. Maria Dholander AA, Almqvist KF, Moens K, Vandekerckhove PJ, Verdonk R, Verdonk P, et al. The use of a prosthetic inlay resurfacing as a salvage procedure for a failed cartilage repair. *Knee Surg Sports Traumatol Arthrosc.* 2015;23(8):2208-12.
 16. Nathwani D, McNicholas M, Hart A, Miles J, Bobić V. Partial resurfacing of the knee with the biopoly implant: Interim report at 2 years. *JB JS Open Access.* 2017;2(2):e0011.
 17. Stålmán A, Sköldenberg O, Martínez-Carranza N, Roberts D, Högstrom M, Ryd L. No implant migration and good subjective outcome of a novel customized femoral resurfacing metal implant for focal chondral lesions. *Knee Surg Sports Traumatol Arthrosc.* 2018;26(7):2196-204.
 18. Beyzadeoglu T, Pehlivanoglu T. Biological response following inlay arthroplasty of the knee: Cartilage flow over the implant. *Cartilage.* 2018;9(2):156-60.
 19. Bollars P, Bosquet M, Vandekerckhove B, Hardeman F, Bellemans J. Prosthetic inlay resurfacing for the treatment of focal, full thickness cartilage defects of the femoral condyle: A bridge between biologics and conventional arthroplasty. *Knee Surg Sports Traumatol Arthrosc.* 2012;20(9):1753-9.
 20. Becher C, Cantiller EB. Focal articular prosthetic resurfacing for the treatment of full-thickness articular cartilage defects in the knee: 12-year follow-up of two cases and review of the literature. *Arch Orthop Trauma Surg.* 2017;137(9):1307-17.
 21. Graves S. Australian Orthopedic Association National Joint Replacement Registry. 2019:190.
 22. Laursen JO, Lind M. Treatment of full-thickness femoral cartilage lesions using condyle resurfacing prosthesis. *Knee Surg Sports Traumatol Arthrosc.* 2017;25(3):746-51.
 23. Ryd L. The mini-metal concept for treating focal lesions and its possible application in athletes. *ASPETAR Sports Med J.* 2016;10:292-5.
 24. Schell H, Zimpfer E, Schmidt-Bleek K, Jung T, Duda GN, Ryd L. Treatment of osteochondral defects: Chondrointegration of metal implants improves after hydroxyapatite coating. *Knee Surg Sports Traumatol Arthrosc.* 2019;27(11):3575-82.
 25. Robertsson O. Annual Report Swedish Knee Arthroplasty Register. 2019.
 26. Sarmiento A. Austin Moore prosthesis in the arthritic hip. Experiences in 224 patients. *Clin Orthop Rel Res.* 1972;82:14-23.
 27. Zhou Z, Yan F, Sha W, Wang L, Zhang X. Unipolar versus bipolar hemiarthroplasty for displaced femoral neck fractures in elderly patients. *Orthopedics.* 2015;38(11):697-702.
 28. Laursen JO. High mid-term revision rate after treatment of large, full-thickness cartilage lesions and OA in the patellofemoral joint using a large inlay resurfacing prosthesis: HemiCAP-Wave(R). *Knee Surg Sports Traumatol Arthrosc.* 2017;25(12):3856-61.
 29. Custers RJH, Creemers LB, van Rijen MHP, Verbout AJ, Saris DBF, Dhert WJA. Cartilage damage caused by metal implants applied for the treatment of established localized cartilage defects in a rabbit model. *J Orthop Res.* 2009;27(1):84-90.
 30. Martínez-Carranza N, Berg HE, Hulthén K, Nurmi-Sandh H, Ryd L, Lagerstedt AS. Focal knee resurfacing and effects of surgical precision on opposing cartilage. A pilot study on 12 sheep. *Osteoarthritis Cartilage.* 2013;21(5):739-45.
 31. Barker MK, Seedhom BB. The relationship of the compressive modulus of articular cartilage with its deformation response to cyclic loading: Does cartilage optimize its modulus so as to minimize the strains arising in it due to the prevalent loading regime? *Rheumatology (Oxford).* 2001;40(3):274-84.
 32. Sanfridsson J, Ryd L, Svahn G, Fridén T, Jonsson K. Radiographic measurement of femorotibial rotation in weight-bearing. The influence of flexion and extension in the knee on the extensor mechanism and angles of the lower extremity in a healthy population. *Acta Radiol.* 2001;42(2):207-17.
 33. Schmalzried TP, Jasty M, Harris WH. Periprosthetic bone loss in total hip arthroplasty. Polyethylene wear debris and the concept of the effective joint space. *J Bone Joint Surg Am.* 1992;74(6):849-63.
 34. Findlay DM. If good things come from above, do bad things come from below? *Arthritis Res Ther.* 2010;12(3):119.
 35. McArthur BA, Scully R, Ross FP, Bostrom MPG, Falghren A. Mechanically induced periprosthetic osteolysis: A systematic review. *HSS J.* 2019;15(3):286-96.
 36. Bergfeld JA. Articular cartilage injury: Filling potholes. *Orthopedics.* 2004;27(9):973-4.
 37. Madry H, Kon E, Condello V, Peretti GM, Steinwachs M, Seil R. Early osteoarthritis of the knee. *Knee Surg Sports Traumatol Arthrosc.* 2016;24(6):1753-62.
 38. Brånemark PI. Osseointegration and its experimental background. *J Prosthet Dent.* 1983;50(3):399-410.
 39. Kremers HM, Lewallen EA, van Wijnen AJ, Lewallen DG. Clinical factors, disease parameters and molecular therapies affecting osseointegration of orthopedic implants. *Curr Moll Biol Rep.* 2016;2(3):123-32.
 40. Martínez-Carranza N, Berg HE, Lagerstedt AS, Nurmi-Sandh H, Schupbach P, Ryd L. Fixation of a double-coated titanium-hydroxyapatite focal knee resurfacing implant: A 12-month Study in sheep. *Osteoarthritis Cartilage.* 2014;22(6):836-44.
 41. Soballe K, Toksvig-Larsen S, Gelineck J, Fruensgaard S, Hansen ES, Ryd L, et al. Migration of hydroxyapatite coated femoral prostheses. A Roentgen Stereophotogrammetric study. *J Bone Joint Surg Br.* 1993;75(5):681-7.
 42. Hogrefe C, Joos H, Maheswaran V, Dürselen L, Ignatius A, Brenne RE.

- Single impact cartilage trauma and TNF-alpha: Interactive effects do not increase early cell death and indicate the need for bi-/multidirectional therapeutic approaches. *Int J Mol Med.* 2012;30(5):1225-32.
43. Vannini F, Spalding S, Andriolo L, Berruto M, Denti M, Espregueira-Mendes J, et al. Sport and early osteoarthritis: The role of sport in aetiology, progression and treatment of knee osteoarthritis. *Knee Surg Sports Traumatol Arthrosc.* 2016;24(6):1786-96.
44. Heinegard D, Saxne T. The role of the cartilage matrix in osteoarthritis. *Nat Rev Rheumatol.* 2011;7(1):50-6.
45. Saris DBF, Dhert WJA, Verbout AJ. Joint homeostasis. The discrepancy between old and fresh defects in cartilage repair. *J Bone Joint Surg Br.* 2003;85(7):1067-76.